

Cepheid C360 Admin v1.3 Software Release Notes

This document contains new information, known limitations, and clarifications not described in the *Cepheid C360 Admin Institution Admin Manual*.

New Features

The latest release includes the following new features or improvements:

- Manually enable/disable users and delete users – Administrators can now manually enable/disable users. Disabled users cannot login to Cepheid C360 Admin. Users that should be permanently disabled can now be deleted so the username can be reused.
- Username login – Users should use their username to login to C360 Admin (previously, email was used for login).
- Lab assistant users are not allowed to login to C360 Admin.
- Configurable period for automatic disablement of inactive user accounts – Administrators can now configure the maximum number of inactive day before a user account is automatically disabled in the Institution Policy. Inactive users are automatically disabled after 365 days if no value is configured.
- Bulk add users – Administrators can now bulk add users by uploading a CSV file.
- Passwords for Lab Assistant users – Institution Administrators can configure passwords for Lab Assistant users.
- Optional simple password – Institution Administrators may now configure optional simple passwords with less password complexity for Lab Assistant users.
- Institution alias – The institution alias field was added to the login page to enforce global uniqueness. The alias is set when the institution is first set up.
- Omni error descriptions – Error descriptions are available in addition to error codes.
- Omni last communication time – Users can view the time the Omni instrument last communicated with C360 Admin in the Mobile Devices tab.
- Mobile devices event history document activity added – The Mobile Devices Event History page now includes the new “document” activity type.
- Browser support – MS Edge is now a supported browser. Internet Explorer 11 (IE 11) is no longer supported.

Known Issues

1. Printers persist in Installation Report

- 1.1. Printers that are removed from the Omni device may display in the Installation Report even after being removed.

2. Emails from C360

- 2.1. Certain email account domains may not receive C360 emails in their inbox. Cepheid recommends checking the junk or spam folders for emails from C360. If C360 emails are flagged by spam filters, Cepheid recommends marking them as not spam.

Cepheid C360 Admin v1.3 Software Release Notes

3. Institution Policy

- 3.1. LIS Configuration and QC Lockout settings in the Institution Policy are not used by the Mobile Device.

4. Audit events

- 4.1. Duplicate audit events will be displayed in UI that are related to a user assigned to multiple laboratories.

5. Logout from C360

- 5.1. Users with certain usernames will logout automatically from C360 Admin. Cepheid recommends to contact Cepheid tech support in case of such issues.